



Quality Control Certificate

Etanercept ELISA (mAb-based) AB102

REFIG-AB102LOTAB102-103	10.07.2026 (dd,mm,yyyy)	
-------------------------	-------------------------	--

Standard Curve			
10X	1:10 Diluted	Measured Value	Range OD
	Study Concentration	OD 450/620nm	
	(ng/mL)		
Standard A	200	3.112	≥ 1.500
Standard B	60	1.508	-
Standard C	20	0.533	-
Standard D	6	0.184	-
Standard E	0	0.055	≤ 0.100

The test results are valid, as long as all criterias are met.

This product has been tested successfully by the Quality Control Department and was released for sale according to existing specifications.

Validation Date:10.07.2024

QC: PASS Ì