



## Quality Control Certificate

### Antibody to Bevacizumab ELISA BB104

<b>REF</b> IG-BB104	<b>LOT</b> BB104-204	 10.10.2026	<b>IVD</b> 
---------------------	----------------------	--	--

Standards/Positive Control/Cut-Off			
	Ready-to-use Concentration (ng/mL)	Measured Value (OD450nm)	Range OD
Standard A	1000	1.265	>0,500
Standard B	300	0.848	>Standard C
Standard C	100	0.396	>Standard D
Standard D	30	0.197	>Standard E
Standard E	0	0.057	<0,100
Positive Control	180 ng/mL ( $\pm$ 20%)	0.585	
Cut-Off (CO)			StE<CO<StD

The test results are valid, as long as all criterias are met.

This product has been tested successfully by the Quality Control Department and was released for sale according to existing specifications.

Validation Date:10.10.2024

**QC: PASS**

