



Quality Control Certificate

Total Antibody to Infliximab ELISA TA101

REF IG-TA101	LOT TA101-105	 13.12.2026 (dd,mm,yy)	IVD 
---------------------	----------------------	---	--

Control Values				
	Measured Value		Range	
	OD	tAU/mL	OD	tAU/mL
Negative Control	0.063		≤ 0.150	
Positive Control	2.430	192	≥ 1.000	

The test results are valid, as long as all criterias are met.

This product has been tested successfully by the Quality Control Department and was released for sale according to existing specifications.

Validation Date:13.12.2024

QC: PASS

