



Quality Control Certificate

Antibody to Rituximab ELISA BB106

REF IG-BB106	LOT BB106-204	 12.03.2027	IVD 
---------------------	----------------------	--	--

Standards/Positive Control/Cut-Off			
	Ready-to-use Concentration (ng/mL)	Measured Value (OD450nm)	Range OD
Standard A	1000	1.494	>0,500
Standard B	300	0.594	>Standard C
Standard C	100	0.254	>Standard D
Standard D	30	0.126	>Standard E
Standard E	0	0.055	<0,100
Positive Control	476 ng/mL(± 20%)	0.766	
Cut-Off (CO)			StE<CO<StD

The test results are valid, as long as all criterias are met.

This product has been tested successfully by the Quality Control Department and was released for sale according to existing specifications.

Validation Date:12.03.2025

QC: PASS

